KING LTS-D COMPETENCY EVALUATION

(This test is to be administered at the end of the King LTS-D training course for documentation of competency in recognition of respiratory emergencies and use of the King LTS-D non-visualized airway device.

EMT Name _____ EMT License # _____

Service Affiliation _____

Performance Criteria	S	U	Comments
Uses appropriate PPE (assume scene safe)			
Verbalize contraindications – 1. Intact gag reflex 2. Known esophageal disease 3. Ingested caustic substances 4. Patient with tracheostomy or laryngectomy 5. Foreign body in trachea 6. Patient under 4 feet			
Check and lubricate cuff as needed (during AED sequence and/or CPR)			
Remove oropharyngeal airway			
With head in appropriate position, applies chin lift			
Inserts airway device, advances tip under base of tongue while rotating tube back to midline. Advance tube until base of connector is aligned with teeth or gums			
Inflates cuffs: Size 3 – 50 ml Size 4 – 70 ml Size 5 – 80 ml			
Removes syringe after inflation of cuffs			
Ventilates while withdrawing tube until ventilation is easy and free flowing			
Confirms placement by observing chest rise and fall and auscultating lungs and stomach (and CO2 monitoring)			
Secures airway in place			
Verbalize indications for removal			
Verbalize proper documentation			

Overall skill performance: _____ Satisfactory _____ Unsatisfactory

Evaluated By _____ Date _____